

CORRIGENDUM

With reference to tender no. 41/GMC/Purchase/Pulse Oximeter/2016-17 the tender specification mentioned in the said tender document are hereby ammended as follows:-

Tender Document Page no. & Clause No.	As in the tender document	Amendment in the tender document
Section IV, Page no. 20	<p>Pulse Oximeter For SNCU (Qty. 8)</p> <ul style="list-style-type: none">• LED Screen• Large display with clear size digit and graph.• Display SpO₂, NIBP and Pulse Rate.• Soft touch key panel for easy operation.• Visual and audible alarm.• Build in rechargeable battery with good back up.	<p>Pulse Oximeter For SNCU (Qty. 8)</p> <ul style="list-style-type: none">• Large LED Size: 4" to 6" only SPo₂ & heart rate display for Numeric & LCD Display for Spo₂ Wave Form.• As Above• Display Spo₂ and pulse Rate• Soft touch key panel for easy operation.• Visual and audible alarm.• Build in rechargeable battery with 10 Hrs back up.• The Pulse Oximeter should be ISO, US FDA & CE neonatal patient use.


Principal
Government Medical College
Haldwani